

Network System Testing and Deployment Plan

Prepared by

Landry Bernard

Steve Stanic

Arne Diercks

Vishwamithra Sunkara

Tuesday, August 3, 2021

Division of Marine Science
School of Ocean Science and Engineering
The University of Southern Mississippi
1020 Balch Blvd. | Stennis Space Center, MS 39529

PURPOSE

This plan provides guidance for initial testing and deployment of the Network System.

TEAM

Team Lead	Steve Stanic, Landry Bernard
Deployment Chief	Arne Diercks
Data	Vishwamithra Sunkara

OUTLINE

- **Conduct Integration Test** of Network System sensor package at MRC Connect entire array of sensors to TRBM and Surface buoy to confirm operational status of all sensors.
 - Test found that 2 of the Temperature sensors were not responding properly during sleep and wake cycles. Those sensors were returned to Manufacturer for replacement.
 - Two new sensors were delivered and tested out fully functional.
- **Acquire data** from all temperature sensors under controlled conditions in a single water bath for pre deployment inspection.
- **Final System performance Evaluation** prior to deployment under controlled condition at MRC
- Securely pack and load entire network system onto deployment vessel
- **Deploy Sensor Array** at a location North of Cat Island in the MS Sound.
- **Acquire data** from Network System via remote Data transfer

TEST OBJECTIVES

INITIAL ACCEPTANCE AND INTEGRATION TEST

- 1) Setup entire Network System at Marine Research Center (MRC) in Gulfport.
- 2) Power on system and monitor output at buoy.
- 3) Connect via remote access to buoy and observe data acquisition and data transfer to server.
- 4) Connect via remote access into the server to monitor data arrival.
- 5) Collect several days' worth of data from system.
- 6) QC incoming data to verify functionality of sensors and acquired data against standard.
- 7) Evaluate and mark erroneous sensor returns.
- 8) Return faulty sensors for repair.


Initial Acceptance and Integration Test Summary

- 1) The entire Network System is setup at MRC in Gulfport.
- 2) All the electrical connections are made as per the electrical schematic provided by the vendor. The system was powered on using the building power supply and monitored the outputs of all sensors for any loose connections.
- 3) Connected a work laptop to the buoy via buoy Wi-Fi and cellular network. Monitored the data acquisition rates of the sensors and the data flow from the sensors to the server.
- 4) Accessed the sensor data from the database and displayed the data in the html based open-source data visualization platform Grafana.
- 5) Noticed that two temperature sensors outputs were intermittent. Troubleshooted these two sensors and found that these two sensors were faulty. Sent faulty sensors to the manufacturer for repairs.
- 6) The Network System was kept powered up collecting several days of data to monitor the health of the system and perform QC of incoming data.

Pre-Deployment Test


- 1) Compare and measure the performance of the 18 RBR RBRCoda³-T sensors to establish measurement baseline.
- 2) Compare and measure the performance of the 2 RBR RBRCoda³-C.T.D. sensors to establish measurement baseline.
- 3) Compare RBRCoda³-T data to RBR RBRCoda³-C.T.D. temperatures to establish baseline and offsets of sensors.
- 4) Immerse all in water temperature sensors in a single container to evaluate performance of entire sensor package.
- 5) Record data output and note any outliers during test, given the calibration data.
- 6) Ready entire network system for deployment, making as few as possible disconnects.

Pre-Deployment Test Summary



Temperature sensors data over a 20 day sampling period, showed a variation of 0.26 °C during in-water test (see Figures to the left). Sensors were allowed to equilibrate for several days submersed in the same water bucket. Strongest variations in measured temperatures between individual sensors occurred during the shallowest slope in of measurements of temperature over time.


Variations of individual sensors of the mean temperature will be calculated to provide individual offsets for each of the sensors.



Stats for 18 individual RBR Temp sensors, including deviation of average mean over all sensor average.


	TR1A Temp 1	TR1A Temp 2	TR1B Temp 1	TR1B Temp 2	TR1C Temp 2	TR2A Temp 1	TR2A Temp 2	TR2B Temp 1	TR2C Temp 1	TR2C Temp 2	TR3A Temp 1	TR3A Temp 2	TR3B Temp 1	TR3B Temp 2	TR3C Temp 2	TR2B Temp 2	TR1C Temp 1	TR3C Temp 1	Average Temp over all sensors
N	1346	1344	1344	1344	1344	1344	1344	1344	1344	1344	1341	1344	1344	1344	1344	1344	1341	1259	
N Missing	2	4	4	4	4	4	4	4	4	4	7	4	4	4	4	4	7	89	
Mean	26.10	26.13	26.14	26.20	26.20	26.11	26.11	26.10	26.10	26.12	26.10	26.13	26.10	26.11	26.10	26.10	26.11	26.07	26.12
Deviation from Average Mean	0.02	-0.02	-0.02	-0.08	-0.08	0.01	0.01	0.02	0.02	0.00	0.02	-0.01	0.02	0.01	0.02	0.02	0.00	0.04	
Min	23.11	23.15	23.16	23.21	23.27	23.17	23.13	23.14	23.12	23.14	23.10	23.13	23.12	23.11	23.10	23.11	23.15	23.15	23.14
Deviation from Average Min	0.03	-0.01	-0.02	-0.06	-0.13	-0.03	0.01	0.00	0.02	0.00	0.04	0.02	0.03	0.04	0.05	0.03	-0.01	-0.01	
Max	28.36	28.40	28.45	28.49	28.47	28.37	28.38	28.37	28.34	28.38	28.42	28.46	28.36	28.39	28.37	28.37	28.36	28.34	28.39
Deviation from Average MAX	0.03	-0.01	-0.05	-0.10	-0.07	0.02	0.01	0.02	0.06	0.02	-0.03	-0.07	0.03	0.01	0.02	0.03	0.03	0.05	
Range	5.25	5.25	5.29	5.28	5.20	5.20	5.25	5.23	5.21	5.24	5.32	5.34	5.25	5.28	5.27	5.25	5.21	5.19	
CV	4.73	4.74	4.74	4.74	4.69	4.72	4.74	4.74	4.69	4.70	4.74	4.79	4.74	4.76	4.76	4.73	4.69	4.71	
Median	26.41	26.45	26.46	26.52	26.50	26.42	26.42	26.41	26.40	26.42	26.41	26.45	26.42	26.42	26.41	26.42	26.42	26.39	
Geometric Mean	26.07	26.11	26.11	26.17	26.17	26.08	26.08	26.07	26.07	26.09	26.07	26.10	26.07	26.08	26.07	26.07	26.09	26.05	
Interquartile Range	1.86	1.86	1.87	1.87	1.86	1.86	1.86	1.86	1.85	1.87	1.86	1.87	1.85	1.86	1.86	1.87	1.86	1.87	
Quantiles90	27.52	27.57	27.57	27.63	27.62	27.53	27.53	27.53	27.52	27.54	27.52	27.55	27.52	27.54	27.53	27.52	27.53	27.48	

CTDA/B and TR2A temp1/2




Matched Pairs

Difference: TR2A Temp 2: temperature-TR2A Temp 1: temperature




TR2A Temp 2: temperature	27.1865	t-Ratio	4.308852
TR2A Temp 1: temperature	27.175	DF	164
Mean Difference	0.01144	Prob > t	<.0001*
Std Error	0.00265	Prob > t	<.0001*
Upper 99%	0.01835	Prob < t	1.0000
Lower 99%	0.00452		
N	165		
Correlation	0.99751		

Difference: CTDA CTD: temperature-TR2A Temp 1: temperature




CTDA CTD: temperature	27.1508	t-Ratio	-10.1957
TR2A Temp 1: temperature	27.175	DF	164
Mean Difference	-0.0242	Prob > t	<.0001*
Std Error	0.00237	Prob > t	1.0000
Upper 99%	-0.018	Prob < t	<.0001*
Lower 99%	-0.0304		
N	165		
Correlation	0.9984		

Difference: CTDB CTD: temperature-TR2A Temp 1: temperature




CTDB CTD: temperature	27.1459	t-Ratio	-12.1214
TR2A Temp 1: temperature	27.175	DF	164
Mean Difference	-0.0292	Prob > t	<.0001*
Std Error	0.00241	Prob > t	1.0000
Upper 99%	-0.0229	Prob < t	<.0001*
Lower 99%	-0.0354		
N	165		
Correlation	0.99833		

Difference: CTDA CTD: temperature-TR2A Temp 2: temperature




CTDA CTD: temperature	27.1508	t-Ratio	-17.9841
TR2A Temp 2: temperature	27.1865	DF	164
Mean Difference	-0.0356	Prob > t	<.0001*
Std Error	0.00198	Prob > t	1.0000
Upper 99%	-0.0305	Prob < t	<.0001*
Lower 99%	-0.0408		
N	165		
Correlation	0.99892		

Difference: CTDB CTD: temperature-TR2A Temp 2: temperature



CTDB CTD: temperature	27.1459	t-Ratio	-20.959
TR2A Temp 2: temperature	27.1865	DF	164
Mean Difference	-0.0406	Prob > t	<.0001*
Std Error	0.00194	Prob > t	1.0000
Upper 99%	-0.0356	Prob < t	<.0001*
Lower 99%	-0.0457		
N	165		
Correlation	0.99897		

Difference: CTDB CTD: temperature-CTDA CTD: temperature



CTDB CTD: temperature	27.1459	t-Ratio	-6.61735
CTDA CTD: temperature	27.1508	DF	164
Mean Difference	-0.005	Prob > t	<.0001*
Std Error	0.00075	Prob > t	1.0000
Upper 99%	-0.003	Prob < t	<.0001*
Lower 99%	-0.0069		
N	165		
Correlation	0.99979		

FIELD INTEGRATION TEST (FIT):

Re-assembled Network System will be deployed in multiple phases

- 1) Phase 1 is the deployment of the TRBM and the surface USM-R2 buoy with its anchors.
- 2) Phase 2 will utilize marker buoys deployed by a small surface vessel at predetermined distances to the location of the TRBM.
- 3) Once the marker buoys are deployed, individual node point anchors with sensor electronics attached to them will be deployed and connections to the node will be established by divers.
- 4) When completely assembled, a final test will be performed to confirm operational status of the Network System.
 - Test Wi-Fi connectivity with surface buoy
 - Test remote download of data via cell network from buoy

PHASE 1 (AFTER DEPLOYMENT OF NEXSENS CB-650 SURFACE DATA BUOY AND TRBM):

Establish buoy communications: The buoy is deployed after powering the system by switching the “Power” button on the buoy to “ON” position. The cellular network takes few (2 to 5) minutes to startup and establish the cellular connection. Then, after the buoy is deployed, a work laptop is connected to the buoy via Wi-Fi and cellular network. Once connected to the buoy, access the modem IP address and enter the username and password. Close and Open the I/O #1 and wait 50 seconds. Access VAK (The buoy Data Acquisition system): Enter <http://<modem ip>:4201>. Once connected to VAK, press the Prevent Shutdown button to prevent it to enter sleep mode. In the Home Page, check the logs for any error messages.

Evaluate data communication functionality: Once the buoy is connected to a work laptop, each sensor is turned on one after another. Live charts menu is used to observe the data stream from the sensor that is turned on. Access the **Live Charts** menu and select the instrument from the **Measurement** dropdown menu. Click **Get Values** to fetch the received data from that particular sensor. Do it for all the sensors and check for any errors.

Evaluate observations data from the buoy: Make connection to the OceanCube server and access the Grafana page. Observe the data transmitted by the buoy to the database on corresponding Grafana dashboards for few data cycles. Check each corresponding dashboard of the sensors equipped on the buoy for any erroneous activity. The buoy is equipped with weather station, wind sensor and tide sensor.

Evaluate real time hydrophone data: The hydrophone data is not visualized on Grafana. To access the hydrophone, use the hydrophone IP address. Turn on the hydrophone and view the live charts on the webpage for any errors. Monitor the hydrophone data for few data cycles.

PHASE 2 (AFTER ALL NODE SENSOR MOORINGS HAVE BEEN DEPLOYED AND CONNECTED TO TRBM):

Evaluate all sensor data in real time: While the whole network system is deployed and operational, monitor the health of the system by checking the observations data on Grafana

dashboards. Check for any intermittent and/or erroneous data reception. Use calibration certificates to check the quality and the measurement accuracy of the sensors.

Evaluate data arrival at shore-based server: Monitor the through-put of the server and database while the network system is operational. Check if the power requirements of each sensor are within the safe limits by monitoring the “instrument Hub” dashboard of the corresponding sensor. Use “data transmitter” dashboard to make sure the data usage is within the bandwidth of the cellular modem.

PERFORMANCE EVALUATION DURING FIRST WEEK AFTER FIELD DEPLOYMENT

- 1) Evaluate weather data from buoy
- 2) Evaluate real time ADCP data
- 3) Evaluate real time hydrophone data
- 4) Evaluate all sensor data in real time
- 5) Evaluate data arrival at shore based server



Figure 1: Assembled Network System during testing phase at the MRC.

SYSTEM COMPONENTS:

NEXSENS CB-650 SURFACE DATA BUOY




Figure 2: NexSens CB-650 Data Buoy Two-Point mooring setup



Figure 3: Surface Data Buoy

Table 1: NexSens CB-650 Buoy Specifications

Hull Dimensions	38 (96.52 cm) outside diameter; 22" (55.88 cm) tall
Tower Dimensions	40" (101.60 cm) tall; 7/8 tubular
Data Well Dimensions	10.3" (26.16 cm) inside diameter; 21.5" (54.61 cm) tall
Weight	215 lbs. (97.52 kg)
Buoyancy	650 lbs. (294.84 kg)
Hull Material	Cross-linked polyethylene foam with polyurea coating & stainless steel deck
Tower /Hardware Material	316 stainless steel
Mooring Attachments	1 or 2 point with 3/4" eyenuts

Data are being transmitted by the buoy every 15 min. and transmitted to server at <http://131.95.7.148:8888>

INTEGRATED SENSORS

Vaisala WINDCAP © Ultrasonic Wind Sensor Series WMT700 (at Instrument Hub Buoy: WMT700)

Recorded Data:

1. wind_speed_avg
2. wind_speed_max
3. wind_speed_min
4. wind_direction_avg
5. wind_direction_min
6. wind_direction_max



Figure 4: Vaisala WINDCAP WMT700

Vaisala Weather Transmitter WXT534 (at Instrument Hub Buoy: WXT534)

Recorded Data:

1. wind_speed_avg
2. wind_speed_max
3. wind_speed_min
4. wind_direction_avg
5. wind_direction_min
6. wind_direction_max
7. air_temperature
8. air_humidity
9. air_pressure
10. rain_accumulation
11. rain_duration
12. rain_intensity
13. hail_accumulation
14. hail_duration
15. hail_intensity



Figure 5: Vaisala Weather Transmitter WXT534

Seaview Wave Sensor SVS-603

Recorded Data:

1. Heading
2. wave_height
3. dominant_period
4. dominant_period_FW
5. dominant_period_TW
6. wave_direction
7. max_wave_height
8. max_wave_period
9. angle_roll
10. angle_pitch
11. date_time



Figure 6: Seaview Wave Sensor SVS-603

Solar Charge Controller Renogy Rover

Recorded Data:

1. charging_status
2. load_power
3. load_current
4. load_voltage
5. solar_power
6. solar_voltage
7. solar_current
8. battery_voltage
9. battery_percentage
10. battery_temperature
11. controller_temperature
12. charging_amp_hours_today
13. discharging_amp_hours_today
14. power_generation_today




Figure 7: Solar Charge Controller Renogy Rover

TRAWL RESISTANT BOTTOM MOUNT (TRBM) BY MOORING SYSTEMS, INC.

General Purpose TRBM Specifications



Figure 8: General Purpose Trawl Resistant Bottom Mount (Mooring Systems Inc.)

Table 2: General Purpose TRBM Specifications

Cover Material	3/8” (9.5 mm) Urethane
Base Material	1” (25.4 mm) Fiberglass
Gimbaled Mount	Molded Urethane Ring
Fasteners	316 Stainless
Length	70” (1775 mm)
Width	50” (1270 mm)
Height (outside)	22” (558 mm)
Height (inside)	19” (480 mm)
Weight in Air	132 lbs. (60 kg) empty
Weight in Water	50 lbs. (23 kg) empty

INSTRUMENT HUB TRBM

Nortek ADCP

Recorded Data:

1. sound_speed
2. heading
3. pitch
4. roll
5. pressure
6. temperature



Figure 9: Nortek ADCP

RBRcoda³ D Tide Sensor (at Instrument Hub TRBM: Tidal Sensor'')

Recorded Data:

1. pressure



Figure 10: RBRcoda³ D Tide Sensor

OCEAN SONICS LTD ICLISTEN HYDROPHONE MODEL RB9-ETH, S/N 6126



Figure 11: ICLISTEN HYDROPHONE

TEMPERATURE NODE SETUP:

A total of 9 small moorings, each with 2 RBRcoda³-T sensors, connected to the central node computer in the TRBM via a single junction box located on the anchor.

Three of these moorings are combined in each of the three arms extending outward from the TRBM Node located at the center (see Figure 6 below).

Table 3:RBRCoda3 T Temperature Sensor

Sensor	S/N	Sample Interval (Seconds)	Sample period (cycles per hour)
TR-A-1-1	205778	10	5
TR-A-1-2	205766	10	5
TR-A-2-1	205768	10	5
TR-A-2-2	205770	10	5
TR-A-3-1	205779	10	5
TR-A-3-2	205764	10	5
TR-B-1-1	205772	10	5
TR-B-1-2	205773	10	5
TR-B-2-1	205769	10	5
TR-B-2-2	205762	10	5
TR-B-3-1	205771	10	5
TR-B-3-2	205777	10	5
TR-C-1-1	207219	10	5
TR-C-1-2	205775	10	5
TR-C-2-1	205767	10	5
TR-C-2-2	205763	10	5
TR-C-3-1	207220	10	5
TR-C-3-2	205765	10	5




Figure 12: Temperature Node.

CTD NODE SETUP:

Two moorings, each equipped with a RBR*concerto*³-C.T.D. multi channel data logger are part of the Network System Setup. CTD-A and CTD-B, each a single mooring with 1 CTD and at the base to the Node-Communications junction box, connected by cable to the TRBM.

Name	Serial Number
CTD-A	205781
CTD-B	205782

Recorded Data:

1. Conductivity
2. Temperature
3. Pressure
4. odo_temperature
5. oxygen_concentration
6. sea_pressure
7. depth
8. salinity
9. speed_of_sound
10. specific_conductivity
11. oxygen_saturation




Figure 13:CTD Node wit RBR CTD.





Figure 14: Planned Network System Deployment Layout after completed Phase 2.

Deployment Plan

M/V Tyson B

August 23 to August 27- 2021

- 1) Network system will be deployed north of Cat Island as part of a mooring deployment and recovery cruise



8/23/2021 MOBILIZATION AFTER 8AM IN GPT HARBOR.

Network-array surface buoy, two concrete anchors (500lbs each), TRBM and
 4 safety buoys with 4 truck brake drum anchors (one drum per buoy).
 Surface Marker buoys to mark node

After loading transit to Cat Island Deployment site

Deployment order:

Surface Buoy	30° 15.65998' N, 89° 4.80392' W
Network Array TRBM	30° 15.65998' N, 89° 4.80392' W
Network Array Safety Buoy 1	30° 15.65998' N, 89° 4.80392' W
Network Array Safety Buoy 2	30° 15.68701' N, 89° 4.77271' W
Network Array Safety Buoy 3	30° 15.63271' N, 89° 4.77271' W
Network Array Safety Buoy 4	30° 15.63294' N, 89° 4.83523' W

4 lifts within reach of crane (20m distance from each other, Tyson B crane has a 39m reach, with 8,000 lbs capacity at full extension)


Suggested order of deployment:

1. concrete anchor #1
2. surface buoy
3. concrete anchor #2
4. TRBM /Node

We will mark the location of TRBM with a surface float to mark its position and to allow the diver an easy descent to the TRBM to connect the cables to the individual sensors.

4 additional lifts of safety moorings at 50m distance to TRBM, 1 truck brake drum anchor per mooring, those can be "man-handled"

During Phase 2 the small anchors with the temp and CTD sensors will be placed and connected by divers to the TRBM



APPENDIX OF CALIBRATION DATA SHEETS

Pressure Calibration Certificate

RBRcoda³ D|tide16 s/n: 205761
 Sensor rating: 20 dbar s/n: L214795
 Nominal accuracy: 0.05%FS (0.01 dbar)
 Reference instrument: Mensor CPC6000 s/n: 612676


Applied pressure, P _{app} (dbar)	Voltage ratio, V	Measured pressure, P _c (dbar)	Calibration error (dbar)	Coefficients
10.027	0.129693	10.0274	-0.0000	C0: -39.147384E-3
12.000	0.154228	11.9999	-0.0001	C1: 79.71752
14.000	0.179067	14.0001	-0.0001	C2: 2.2275455
16.000	0.203864	16.0001	0.0001	C3: 758.9983E-3
18.000	0.228615	17.9996	-0.0001	X0: 10.0296
20.000	0.253333	20.0000	0.0001	X1: 7.623532E-3
22.000	0.278007	21.9999	-0.0001	X2: 66.76267E-6
24.000	0.302637	24.0001	0.0000	X3: -290.5105E-9
26.000	0.327224	26.0001	-0.0000	X4: 322.2121E-6
28.000	0.351765	28.0000	-0.0001	X5: 22.099691
30.000	0.376263	29.9999	-0.0001	

$$P_c = X_0 + \frac{P_m - X_0 - X_1(T - X_5) - X_2(T - X_5)^2 - X_3(T - X_5)^3}{1 + X_4(T - X_5)}$$


Head (mm) = 311

$$P_m = C_0 + C_1V + C_2V^2 + C_3V^3$$

Calibration error vs. Pressure (Tcal = 22.1°C)



Calibration error vs. Temperature (Patm = 10.05 dbar)



Calibration Date: 2020-11-28
 Issue Date: 2020-11-30
 File Name: 205761_20201130_1120P.rsk

Operator: *Duong*
 dluong


Approver: *Kmalorny*
 kmalorny



Temperature Calibration Certificate

Logger ID: RBRcoda³ Serial No: 205762 Channel No: 1

Reference Temperature, ITS-90	Voltage ratio, V	Measured Temperature, ITS-90	Calibration error	Coefficients
-4.97140	0.739412	-4.97143	-0.00003	C0: 3.464633E-3
0.67275	0.677400	0.67287	0.00011	C1: -250.66652E-6
6.89799	0.603917	6.89784	-0.00014	C2: 2.508184E-6
12.80487	0.532249	12.80491	0.00005	C3: -64.72184E-9
18.57324	0.463383	18.57327	0.00003	
24.31829	0.398359	24.31829	-0.00000	
30.11210	0.338212	30.11207	-0.00002	
34.99716	0.292554	34.99717	0.00001	



Calibration Date: 2020-12-04
Issue Date: 2020-12-05
Calibration ID: 42925

Operator: T. Akwethel
takuetteh


Approver: [Signature]
kmalorny



Temperature Calibration Certificate

Logger ID: RBRcoda³ Serial No: 205763 Channel No: 1

Reference Temperature, ITS-90	Voltage ratio, V	Measured Temperature, ITS-90	Calibration error	Coefficients	
-5.04009	0.714294	-5.04010	-0.00001	C0:	3.4952562E-3
0.61087	0.649622	0.61087	0.00000	C1:	-253.69474E-6
6.84389	0.574366	6.84394	0.00005	C2:	2.4322385E-6
12.76131	0.502298	12.76129	-0.00002	C3:	-61.200296E-9
18.54418	0.434186	18.54404	-0.00013		
24.30144	0.370907	24.30158	0.00014		
30.10570	0.313219	30.10569	-0.00001		
35.00054	0.269917	35.00052	-0.00002		



Calibration Date: 2020-12-01
Issue Date: 2020-12-03
Calibration ID: 42846

Operator: *Duong*
duong

Approver: *Kmalorny*
kmalorny




Temperature Calibration Certificate

Logger ID: RBRcoda³ Serial No: 205764 Channel No: 1

Reference Temperature, ITS-90	Voltage ratio, V	Measured Temperature, ITS-90	Calibration error	Coefficients	
-4.95948	0.721911	-4.95944	0.00003	C0:	3.4846377E-3
0.03846	0.665777	0.03845	-0.00001	C1:	-253.43886E-6
6.03589	0.594387	6.03574	-0.00016	C2:	2.4425478E-6
12.03283	0.521556	12.03291	0.00008	C3:	-69.63962E-9
18.02303	0.450492	18.02341	0.00038		
24.02040	0.383698	24.01990	-0.00049		
30.02043	0.323066	30.02059	0.00016		
35.01691	0.278061	35.01692	0.00001		

Residuals vs. Temperature



Calibration Date: 2020-11-29
Issue Date: 2020-11-30
Calibration ID: 42770

Operator: *Duong*
dluong

Approver: *Kmalorny*
kmalorny




Temperature Calibration Certificate

Logger ID: RBRcoda³ Serial No: 205765 Channel No: 1

Reference Temperature, ITS-90	Voltage ratio, V	Measured Temperature, ITS-90	Calibration error	Coefficients
-4.97162	0.705818	-4.97163	-0.00001	C0: 3.5065676E-3
0.67277	0.639777	0.67281	0.00004	C1: -251.66976E-6
6.89799	0.563314	6.89791	-0.00008	C2: 2.5902216E-6
12.80492	0.490547	12.80498	0.00006	C3: -87.77939E-9
18.57309	0.422248	18.57310	0.00001	
24.31798	0.359167	24.31797	-0.00001	
30.11181	0.301995	30.11177	-0.00003	
34.99659	0.259334	34.99661	0.00002	

Residuals vs. Temperature



Calibration Date: 2020-11-29
Issue Date: 2020-12-01
Calibration ID: 42818

Operator: *Duong*
duong


Approver: *Kmalorny*
kmalorny

RBR

Temperature Calibration Certificate

Logger ID: RBRcoda³ Serial No: 205766 Channel No: 1

Reference Temperature, ITS-90	Voltage ratio, V	Measured Temperature, ITS-90	Calibration error	Coefficients
-5.04040	0.691465	-5.04040	-0.00000	C0: 3.523364E-3
0.61067	0.624225	0.61068	0.00001	C1: -253.83904E-6
6.84380	0.547189	6.84376	-0.00004	C2: 2.4265228E-6
12.76129	0.474611	12.76132	0.00003	C3: -59.18108E-9
18.54418	0.407080	18.54423	0.00005	
24.30170	0.345252	24.30160	-0.00011	
30.10624	0.289635	30.10632	0.00007	
35.00115	0.248358	35.00113	-0.00002	



Calibration Date: 2020-11-30
Issue Date: 2020-12-01
Calibration ID: 42809

Operator: *Duong*
dluong

Approver: *Kmalorny*
kmalorny


RBR Limited, 359 Terry Fox Drive, Ottawa ON, K2K 2E7, Canada | +1.613.599.8900 | www.rbr-global.com

RBR

Temperature Calibration Certificate

Logger ID: RBRcoda³ Serial No: 205767 Channel No: 1

Reference Temperature, ITS-90	Voltage ratio, V	Measured Temperature, ITS-90	Calibration error	Coefficients
-5.03906	0.732612	-5.03901	0.00005	C0: 3.4717575E-3
0.61168	0.670254	0.61152	-0.00017	C1: -253.4823E-6
6.84446	0.596760	6.84464	0.00018	C2: 2.437434E-6
12.76180	0.525424	12.76172	-0.00008	C3: -74.36366E-9
18.54426	0.457122	18.54425	-0.00001	
24.30149	0.392880	24.30159	0.00010	
30.10544	0.333643	30.10528	-0.00016	
35.00035	0.288741	35.00042	0.00008	



Calibration Date: 2020-12-02
Issue Date: 2020-12-04
Calibration ID: 42877

Operator: *Duong*
dluong

Approver: *Kmalorny*
kmalorny


RBR Limited, 359 Terry Fox Drive, Ottawa ON, K2K 2E7, Canada | +1.613.599.8900 | www.rbr-global.com



Temperature Calibration Certificate

Logger ID: RBRcoda³ Serial No: 205768 Channel No: 1

Reference Temperature, ITS-90	Voltage ratio, V	Measured Temperature, ITS-90	Calibration error	Coefficients
-5.04009	0.740572	-5.04007	0.00002	C0: 3.4657335E-3
0.61087	0.678190	0.61080	-0.00007	C1: -249.09156E-6
6.84389	0.604162	6.84400	0.00012	C2: 2.4666078E-6
12.76131	0.531887	12.76127	-0.00003	C3: -71.95062E-9
18.54417	0.462394	18.54411	-0.00007	
24.30145	0.396854	24.30146	0.00002	
30.10570	0.336332	30.10576	0.00006	
35.00055	0.290455	35.00051	-0.00003	



Calibration Date: 2020-12-01
Issue Date: 2020-12-03
Calibration ID: 42845

Operator: *Duong*
dluong

Approver: *Kmalorny*
kmalorny




Temperature Calibration Certificate

Logger ID: RBRcoda³ Serial No: 205769 Channel No: 1

Reference Temperature, ITS-90	Voltage ratio, V	Measured Temperature, ITS-90	Calibration error	Coefficients	
-5.04009	0.744078	-5.04007	0.00003	C0:	3.4605055E-3
0.61087	0.682341	0.61080	-0.00007	C1:	-249.61677E-6
6.84389	0.608908	6.84392	0.00003	C2:	2.4619599E-6
12.76131	0.537017	12.76136	0.00006	C3:	-77.07981E-9
18.54418	0.467698	18.54411	-0.00007		
24.30144	0.402132	24.30152	0.00008		
30.10570	0.341414	30.10559	-0.00011		
35.00054	0.295267	35.00059	0.00005		

Residuals vs. Temperature



Calibration Date: 2020-12-01
 Issue Date: 2020-12-02
 Calibration ID: 42843

Operator: *Duong*
 dluong

Approver: *Kmalorny*
 kmalorny




Temperature Calibration Certificate

Logger ID: RBRcoda³ Serial No: 205770 Channel No: 1


Reference Temperature, ITS-90	Voltage ratio, V	Measured Temperature, ITS-90	Calibration error	Coefficients
-4.95948	0.673137	-4.95949	-0.00001	C0: 3.5446994E-3
0.03846	0.612036	0.03851	0.00004	C1: -252.7399E-6
6.03589	0.536619	6.03583	-0.00007	C2: 2.647061E-6
12.03283	0.462249	12.03288	0.00005	C3: -81.622694E-9
18.02303	0.392107	18.02308	0.00006	
24.02040	0.328284	24.02020	-0.00019	
30.02043	0.272086	30.02062	0.00020	
35.01691	0.231413	35.01684	-0.00007	

Residuals vs. Temperature



Calibration Date: 2020-11-29
Issue Date: 2020-11-30
Calibration ID: 42771

Operator: 
dluong

Approver: 
kmalorny




Temperature Calibration Certificate

Logger ID: RBRcoda³ Serial No: 205771 Channel No: 1


Reference Temperature, ITS-90	Voltage ratio, V	Measured Temperature, ITS-90	Calibration error	Coefficients	
-5.03906	0.745803	-5.03896	0.00011	C0:	3.4575902E-3
0.61168	0.684488	0.61143	-0.00025	C1:	-250.03048E-6
6.84446	0.611454	6.84452	0.00006	C2:	2.5742302E-6
12.76180	0.539835	12.76191	0.00011	C3:	-83.24676E-9
18.54426	0.470657	18.54437	0.00011		
24.30150	0.405103	24.30152	0.00002		
30.10544	0.344286	30.10505	-0.00039		
35.00035	0.297979	35.00058	0.00023		

Residuals vs. Temperature



Calibration Date: 2020-12-02
Issue Date: 2020-12-04
Calibration ID: 42878

Operator: 
dluong

Approver: 
kmalorny




Temperature Calibration Certificate

Logger ID: RBRcoda³ Serial No: 205772 Channel No: 1

Reference Temperature, ITS-90	Voltage ratio, V	Measured Temperature, ITS-90	Calibration error	Coefficients
-4.97893	0.706033	-4.97888	0.00004	C0: 3.5078009E-3
0.66288	0.639595	0.66286	-0.00002	C1: -250.14993E-6
6.88391	0.562669	6.88366	-0.00025	C2: 2.5093887E-6
12.78907	0.489448	12.78934	0.00027	C3: -82.14832E-9
18.55706	0.420762	18.55723	0.00017	
24.30274	0.357388	24.30260	-0.00014	
30.10055	0.300004	30.10031	-0.00024	
34.99010	0.257241	34.99027	0.00017	

Residuals vs. Temperature



Calibration Date: 2020-12-04
Issue Date: 2020-12-05
Calibration ID: 42922

Operator: *Duong*
dluong

Approver: *Kmalorny*
kmalorny




Temperature Calibration Certificate

Logger ID: RBRcoda³ Serial No: 205773 Channel No: 1

Reference Temperature, ITS-90	Voltage ratio, V	Measured Temperature, ITS-90	Calibration error	Coefficients	
-4.97162	0.693558	-4.97161	0.00001	C0:	3.5199283E-3
0.67278	0.626649	0.67278	0.00000	C1:	-253.72457E-6
6.89800	0.549878	6.89791	-0.00009	C2:	2.4669785E-6
12.80491	0.477457	12.80497	0.00006	C3:	-79.008785E-9
18.57306	0.410003	18.57315	0.00009		
24.31798	0.348112	24.31798	0.00001		
30.11180	0.292328	30.11162	-0.00018		
34.99660	0.250870	34.99670	0.00010		

Residuals vs. Temperature



Calibration Date: 2020-11-29
Issue Date: 2020-12-01
Calibration ID: 42819

Operator: *Duong*
dluong

Approver: *Kmalorny*
kmalorny




Temperature Calibration Certificate

Logger ID: RBRcoda³ Serial No: 205774 Channel No: 1

Reference Temperature, ITS-90	Voltage ratio, V	Measured Temperature, ITS-90	Calibration error	Coefficients	
-4.97162	0.707877	-4.97167	-0.00005	C0:	3.5029673E-3
0.67278	0.642375	0.67293	0.00015	C1:	-252.98411E-6
6.89800	0.566479	6.89789	-0.00011	C2:	2.4501758E-6
12.80492	0.494158	12.80479	-0.00013	C3:	-86.66383E-9
18.57307	0.426156	18.57326	0.00019		
24.31798	0.363231	24.31796	-0.00001		
30.11180	0.306072	30.11174	-0.00006		
34.99660	0.263326	34.99662	0.00002		

Residuals vs. Temperature



Calibration Date: 2020-11-29
Issue Date: 2020-12-01
Calibration ID: 42817

Operator: Jeff Walker
jwalker


Approver: [Signature]
kmalorny

RBR

Temperature Calibration Certificate

Logger ID: RBRcoda³ Serial No: 205775 Channel No: 1

Reference Temperature, ITS-90	Voltage ratio, V	Measured Temperature, ITS-90	Calibration error	Coefficients
-5.04040	0.750721	-5.04040	-0.00001	C0: 3.4533427E-3
0.61067	0.689554	0.61070	0.00003	C1: -247.91146E-6
6.84381	0.616408	6.84376	-0.00005	C2: 2.5102906E-6
12.76129	0.544421	12.76127	-0.00003	C3: -82.994326E-9
18.54418	0.474688	18.54431	0.00013	
24.30169	0.408479	24.30162	-0.00007	
30.10624	0.346962	30.10621	-0.00003	
35.00115	0.300104	35.00118	0.00003	



Calibration Date: 2020-11-30
Issue Date: 2020-12-01
Calibration ID: 42812

Operator: *Duong*
duong

Approver: *Kmalorny*
kmalorny

RBR Limited, 359 Terry Fox Drive, Ottawa ON, K2K 2E7, Canada | +1.613.599.8900 | www.rbr-global.com




Temperature Calibration Certificate

Logger ID: RBRcoda³ Serial No: 205776 Channel No: 1

Reference Temperature, ITS-90	Voltage ratio, V	Measured Temperature, ITS-90	Calibration error	Coefficients	
-5.04009	0.734829	-5.04005	0.00005	C0:	3.4689445E-3
0.61087	0.672743	0.61074	-0.00014	C1:	-253.38135E-6
6.84389	0.599454	6.84402	0.00013	C2:	2.4242124E-6
12.76131	0.528202	12.76125	-0.00005	C3:	-82.56047E-9
18.54418	0.459871	18.54423	0.00005		
24.30144	0.395515	24.30143	-0.00002		
30.10570	0.336090	30.10563	-0.00007		
35.00054	0.291002	35.00059	0.00005		

Residuals vs. Temperature



Calibration Date: 2020-12-01
Issue Date: 2020-12-03
Calibration ID: 42844

Operator: *Duong*
dluong


Approver: *Kmalorny*
kmalorny



Temperature Calibration Certificate

Logger ID: RBRcoda³ Serial No: 205777 Channel No: 1

Reference Temperature, ITS-90	Voltage ratio, V	Measured Temperature, ITS-90	Calibration error	Coefficients
-5.04039	0.677952	-5.04036	0.00004	C0: 3.5388966E-3
0.61067	0.609576	0.61060	-0.00007	C1: -254.60086E-6
6.84381	0.531939	6.84376	-0.00005	C2: 2.4745598E-6
12.76129	0.459462	12.76138	0.00009	C3: -74.0124E-9
18.54418	0.392592	18.54429	0.00011	
24.30171	0.331837	24.30159	-0.00012	
30.10624	0.277557	30.10618	-0.00006	
35.00115	0.237492	35.00121	0.00007	



Calibration Date: 2020-11-30
Issue Date: 2020-12-01
Calibration ID: 42808

Operator: *Duong*
dluong

Approver: *Kmalorny*
kmalorny


RBR Limited, 359 Terry Fox Drive, Ottawa ON, K2K 2E7, Canada | +1.613.599.8900 | www.rbr-global.com



Temperature Calibration Certificate

Logger ID: RBRcoda³ Serial No: 205778 Channel No: 1

Reference Temperature, ITS-90	Voltage ratio, V	Measured Temperature, ITS-90	Calibration error	Coefficients
-4.97140	0.715467	-4.97141	-0.00000	C0: 3.494863E-3
0.67275	0.650486	0.67274	-0.00001	C1: -251.34353E-6
6.89799	0.574749	6.89807	0.00009	C2: 2.5636348E-6
12.80488	0.502182	12.80477	-0.00011	C3: -70.948616E-9
18.57325	0.433615	18.57323	-0.00002	
24.31829	0.369899	24.31838	0.00010	
30.11210	0.311832	30.11205	-0.00004	
34.99716	0.268296	34.99716	0.00000	



Calibration Date: 2020-12-04
Issue Date: 2020-12-05
Calibration ID: 42926

Operator: T. Akwethel
takuetteh

Approver: [Signature]
kmalorny




Temperature Calibration Certificate

Logger ID: RBRcoda³ Serial No: 205779 Channel No: 1

Reference Temperature, ITS-90	Voltage ratio, V	Measured Temperature, ITS-90	Calibration error	Coefficients	
-5.04040	0.713517	-5.04032	0.00007	C0:	3.4959305E-3
0.61067	0.648856	0.61050	-0.00017	C1:	-253.97725E-6
6.84380	0.573637	6.84383	0.00003	C2:	2.4689934E-6
12.76129	0.501631	12.76138	0.00009	C3:	-91.898855E-9
18.54418	0.433599	18.54431	0.00013		
24.30170	0.370415	24.30154	-0.00015		
30.10624	0.312820	30.10615	-0.00010		
35.00115	0.269596	35.00125	0.00010		

Residuals vs. Temperature



Calibration Date: 2020-11-30
Issue Date: 2020-12-01
Calibration ID: 42810

Operator: *Duong*
dluong

Approver: *Kmalorny*
kmalorny



Conductivity Calibration Certificate

RBRconcerto³ C.T.D.ODO s/n: 205781

References: Autosal8400B#66289, MS-315#15506, SSW P162, RC#002


Reference Resistance (ohm)	Reference Conductivity (mS/cm)	Voltage Ratio, V	Measured Conductivity (mS/cm)	Calibration Error (mS/cm)	Coefficients
open	0.0000	-0.000147	-0.0002	-0.0002	
694.023	6.1588	0.039084	6.1582	-0.0006	C1: 156.97823
331.918	12.8777	0.081886	12.8772	-0.0006	X0: 373.78885E-6
150.011	28.4936	0.181372	28.4944	0.0008	X1: -7.940543E-6
100.007	42.7405	0.272132	42.7416	0.0010	X2: 600E-9
75.013	56.9815	0.362847	56.9820	0.0005	X3: 15.004712
55.511	77.0001	0.490368	77.0000	-0.0001	X4: 10
47.018	90.9088	0.578967	90.9080	-0.0009	

Bath	Voltage Ratio	Temperature (ITS-90)	Salinity (PSS-78)	Conductivity (mS/cm)
T15S35	0.2733414	15.00471	35.0085	42.9315
T25S35	0.3303040	23.86290	34.9901	51.8737


Cell Constant @T15S35 = 4.27435 1/cm

$$C_{cor} = \frac{C_0 + C_1 * V - X_0 * (T - X_3)}{1 + X_1 * (T - X_3) + X_2 * (P - X_4)}$$

Calibration error vs. Conductivity



Calibration error vs. Temperature



Calibration Date: 2020-12-04
 Issue Date: 2020-12-04
 File Name: 205781_20201204_1720C.rsk

Operator: Jeff Walker
 jwalker

Approver: [Signature]
 kmalorny




Temperature Calibration Certificate

Logger ID: RBRconcerto³ Serial No: 205781 Channel No: 2

Reference Temperature, ITS-90	Voltage ratio, V	Measured Temperature, ITS-90	Calibration error	Coefficients	
-5.04009	0.695860	-5.04004	0.00005	C0:	3.5174747E-3
0.61087	0.629284	0.61074	-0.00014	C1:	-254.46268E-6
6.84389	0.552776	6.84394	0.00006	C2:	2.480297E-6
12.76131	0.480456	12.76144	0.00014	C3:	-53.564086E-9
18.54417	0.412937	18.54402	-0.00016		
24.30144	0.350907	24.30155	0.00011		
30.10570	0.294934	30.10558	-0.00012		
35.00054	0.253268	35.00060	0.00006		

Residuals vs. Temperature



Calibration Date: 2020-12-01
Issue Date: 2020-12-02
Calibration ID: 42837

Operator: T. Akwethel
takwethel

Approver: [Signature]
kmalorny



Pressure Calibration Certificate

RBRconcerto³ C.T.D.ODO s/n: 205781

Sensor rating: 20 dbar s/n: L214796

Nominal accuracy: 0.05%FS (0.01 dbar)

Reference instrument: Mensor CPC6000 s/n: 612676


Applied pressure, P _{app} (dbar)	Voltage ratio, V	Measured pressure, P _c (dbar)	Calibration error (dbar)	Coefficients
10.024	0.129494	10.0240	-0.0004	C0: -112.48907E-3
12.000	0.153844	11.9999	-0.0002	C1: 80.49681
14.000	0.178457	13.9998	-0.0001	C2: 2.1269863
16.000	0.203034	15.9998	-0.0002	C3: 608.57624E-3
18.000	0.227577	18.0000	-0.0001	X0: 10.0273
20.000	0.252078	19.9999	-0.0001	X1: 6.897866E-3
22.000	0.276545	22.0000	-0.0000	X2: 69.23961E-6
24.000	0.300971	23.9999	-0.0001	X3: 284.70654E-9
26.000	0.325360	26.0001	-0.0000	X4: 330.06948E-6
28.000	0.349708	28.0001	-0.0000	X5: 21.406427
30.000	0.374014	29.9999	0.0000	

$$P_c = X_0 + \frac{P_m - X_0 - X_1(T - X_5) - X_2(T - X_5)^2 - X_3(T - X_5)^3}{1 + X_4(T - X_5)}$$


Head (mm) = 323

$$P_m = C_0 + C_1V + C_2V^2 + C_3V^3$$

Calibration error vs. Pressure (Tcal = 21.4°C)



Calibration error vs. Temperature (Patm = 9.90 dbar)



Calibration Date: 2020-12-03
 Issue Date: 2020-12-03
 File Name: 205781_20201203_1645P.rsk

Operator: T. Akuetteh
 takuetteh

Approver: [Signature]
 kmalorny




Temperature Calibration Certificate

Logger ID: RBRcoda Serial No: 202309 Channel No: 1


Reference Temperature, ITS-90	Voltage ratio, V	Measured Temperature, ITS-90	Calibration error	Coefficients	
-5.03651	0.699107	-5.03654	-0.00003	C0:	3.5139888E-3
0.61375	0.632729	0.61380	0.00005	C1:	-253.76776E-6
6.84563	0.556272	6.84575	0.00013	C2:	2.5431632E-6
12.76296	0.483823	12.76256	-0.00040	C3:	-48.16859E-9
18.54566	0.416028	18.54597	0.00031		
24.30290	0.353654	24.30290	0.00000		
30.10683	0.297287	30.10672	-0.00010		
35.00142	0.255286	35.00146	0.00003		

Residuals vs. Temperature



Calibration Date: 12/Aug/2019
Issue Date: 13/Aug/2019
Calibration ID: 34434

Operator: 
dluong

Approver: 
kmalorny

RBR Limited, 95 Hines Road, Ottawa, Canada K2K 2M5 | +1 (613) 599-8900 | www.rbr-global.com



Optical DO Calibration Certificate

RBRcoda ODO s/n: 202309

Foil batch: 160530-001

Salinity: 0 PSU

Temperature range: 0 - 35 °C


C0:	0.0	C12:	8.03775
C1:	1.0	C13:	-79.12303E-3
C2:	0.0	C14:	287.52E-6
C3:	1.0	C15:	0.0
C4:	5.296227E3	C16:	-94.67624E-3
C5:	-127.5013	C17:	429.7717E-6
C6:	1.672486	C18:	0.0
C7:	-10.7012E-3	C19:	0.0
C8:	-321.7648	C20:	426.7697E-6
C9:	5.298149	C21:	0.0
C10:	-44.59722E-3	C22:	0.0
C11:	180.123E-6	C23:	0.0

Calibration Date: 2020-04-30


Issue Date: 2020-04-30

Calibration ID: C-078821

Operator:


takwetteh

Approver:


kmalorny

RBR Limited, 95 Hines Road, Ottawa, Canada K2K 2M5 | +1 (613) 599-8900 | www.rbr-global.com



Conductivity Calibration Certificate

RBRconcerto³ C.T.D.ODO s/n: 205782

References: Autosal8400B#66289, MS-315#15506, SSW P162, RC#002


Reference Resistance (ohm)	Reference Conductivity (mS/cm)	Voltage Ratio, V	Measured Conductivity (mS/cm)	Calibration Error (mS/cm)	Coefficients
open	0.0000	-0.000134	-0.0005	-0.0005	C0: 20.529384E-3
694.023	6.1611	0.039109	6.1603	-0.0007	C1: 156.99046
331.918	12.8825	0.081925	12.8820	-0.0005	X0: 272.71733E-6
150.011	28.5041	0.181439	28.5047	0.0007	X1: -12.285235E-6
100.007	42.7563	0.272227	42.7576	0.0014	X2: 600E-9
75.013	57.0025	0.362972	57.0036	0.0012	X3: 15.008923
55.511	77.0284	0.490527	77.0285	0.0001	X4: 10
47.018	90.9423	0.579145	90.9407	-0.0016	

Bath	Voltage Ratio	Temperature (ITS-90)	Salinity (PSS-78)	Conductivity (mS/cm)
T15S35	0.2733613	15.00892	35.0085	42.9356
T25S35	0.3302891	23.86382	34.9911	51.8760


Cell Constant @T15S35 = 4.27593 1/cm

$$C_{cor} = \frac{C_0 + C_1 * V - X_0 * (T - X_3)}{1 + X_1 * (T - X_3) + X_2 * (P - X_4)}$$

Calibration error vs. Conductivity



Calibration error vs. Temperature



Calibration Date: 2020-12-04
 Issue Date: 2020-12-04
 File Name: 205782_20201204_1732C.rsk

Operator: Jeff Walker
 jwalker

Approver: [Signature]
 kmalorny




Temperature Calibration Certificate

Logger ID: RBRconcerto³ Serial No: 205782 Channel No: 2

Reference Temperature, ITS-90	Voltage ratio, V	Measured Temperature, ITS-90	Calibration error	Coefficients	
-4.95942	0.693651	-4.95934	0.00008	C0:	3.5191863E-3
0.03873	0.634739	0.03860	-0.00013	C1:	-254.29667E-6
6.03616	0.561141	6.03607	-0.00010	C2:	2.4613857E-6
12.03305	0.487543	12.03314	0.00009	C3:	-70.66773E-9
18.02350	0.417124	18.02393	0.00043		
24.02055	0.352160	24.02006	-0.00049		
30.02118	0.294175	30.02122	0.00004		
35.01774	0.251731	35.01782	0.00008		

Residuals vs. Temperature



Calibration Date: 2020-11-30
 Issue Date: 2020-12-01
 Calibration ID: 42778

Operator: Jeff Walker
 jwalker

Approver: [Signature]
 kmalorny



Pressure Calibration Certificate

RBRconcerto³ C.T.D.ODO s/n: 205782

Sensor rating: 20 dbar s/n: L214788

Nominal accuracy: 0.05%FS (0.01 dbar)

Reference instrument: Mensor CPC6050 s/n: 41000CAM


Applied pressure, P _{app} (dbar)	Voltage ratio, V	Measured pressure, P _c (dbar)	Calibration error (dbar)	Coefficients
9.962	0.126295	9.9618	0.0002	C0: 8.427022E-3
12.000	0.151216	12.0001	-0.0003	C1: 81.16047
14.000	0.175625	13.9994	-0.0001	C2: 2.1374688
15.999	0.200006	15.9994	0.0001	C3: 637.3461E-3
18.000	0.224353	17.9996	0.0001	X0: 9.9658
20.000	0.248660	19.9996	0.0001	X1: 7.0297155E-3
22.000	0.272941	22.0004	0.0000	X2: 63.714535E-6
24.000	0.297164	23.9996	0.0001	X3: 107.89746E-9
26.000	0.321356	25.9995	-0.0004	X4: 329.78624E-6
27.999	0.345508	27.9992	-0.0001	X5: 22.352448
29.999	0.369625	29.9994	0.0001	

$$P_c = X_0 + \frac{P_m - X_0 - X_1(T - X_5) - X_2(T - X_5)^2 - X_3(T - X_5)^3}{1 + X_4(T - X_5)}$$


Head (mm) = 332

$$P_m = C_0 + C_1V + C_2V^2 + C_3V^3$$

Calibration error vs. Pressure (Tcal = 22.4°C)



Calibration error vs. Temperature (Patm = 9.87 dbar)



Calibration Date: 2020-12-02
 Issue Date: 2020-12-02
 File Name: 205782_20201202_1259P.rsk

Operator: T. Akuehleh
 takuehleh

Approver: [Signature]
 kmalorny




Temperature Calibration Certificate

Logger ID: RBRcoda Serial No: 204593 Channel No: 1

Reference Temperature, ITS-90	Voltage ratio, V	Measured Temperature, ITS-90	Calibration error	Coefficients	
-5.03843	0.728118	-5.03853	-0.00009	C0:	3.4778328E-3
0.61194	0.665108	0.61212	0.00018	C1:	-253.30816E-6
6.84426	0.591087	6.84437	0.00012	C2:	2.4339122E-6
12.76153	0.519475	12.76115	-0.00038	C3:	-68.051904E-9
18.54419	0.451128	18.54418	-0.00001		
24.30163	0.387051	24.30185	0.00022		
30.10608	0.328147	30.10615	0.00007		
35.00061	0.283627	35.00050	-0.00011		

Residuals vs. Temperature



Calibration Date: 2020-06-11
Issue Date: 2020-06-12
Calibration ID: 39926

Operator: *Duong*
dluong

Approver: *Kmalorny*
kmalorny

RBR

Optical DO Calibration Certificate

RBRcoda ODO s/n: 204593
Foil batch: 181002-101_PSt3-1111-02
Salinity: 0 PSU
Temperature range: 0 - 35 °C


C0:	0.0	C12:	8.0462
C1:	1.0	C13:	-82.12904E-3
C2:	0.0	C14:	313.9571E-6
C3:	1.0	C15:	0.0
C4:	4.824596E3	C16:	-98.98348E-3
C5:	-117.6729	C17:	468.9997E-6
C6:	1.56718	C18:	0.0
C7:	-10.22726E-3	C19:	0.0
C8:	-307.3719	C20:	465.5301E-6
C9:	5.184242	C21:	0.0
C10:	-44.91426E-3	C22:	0.0
C11:	186.9222E-6	C23:	0.0

Calibration Date: 2020-07-15
Issue Date: 2020-07-15
Calibration ID: C-083674

Operator:


takuetteh

Approver:


kmalorny

RBR Limited, 95 Hines Road, Ottawa, Canada K2K 2M5 | +1 (613) 599-8900 | www.rbr-global.com




Temperature Calibration Certificate

Logger ID: RBRcoda³ Serial No: 207219 Channel No: 1


Reference Temperature, ITS-90	Voltage ratio, V	Measured Temperature, ITS-90	Calibration error	Coefficients
-4.97476	0.678831	-4.97478	-0.00002	C0: 3.5381648E-3
0.66627	0.610259	0.66634	0.00007	C1: -252.86997E-6
6.88686	0.532334	6.88682	-0.00004	C2: 2.6041919E-6
12.79259	0.459529	12.79239	-0.00021	C3: -79.656296E-9
18.56003	0.392367	18.56035	0.00032	
24.30631	0.331309	24.30631	0.00000	
30.10419	0.276739	30.10395	-0.00024	
34.99336	0.236487	34.99348	0.00012	

Residuals vs. Temperature



Calibration Date: 2021-05-14
Issue Date: 2021-05-17
Calibration ID: 46631

Operator: 
dluong

Approver: 
kmalorny




Temperature Calibration Certificate

Logger ID: RBRcoda³ Serial No: 207220 Channel No: 1


Reference Temperature, ITS-90	Voltage ratio, V	Measured Temperature, ITS-90	Calibration error	Coefficients
-4.97476	0.693416	-4.97475	0.00001	C0: 3.5195563E-3
0.66626	0.626729	0.66627	0.00002	C1: -254.43753E-6
6.88686	0.550237	6.88678	-0.00009	C2: 2.475926E-6
12.79258	0.478036	12.79248	-0.00010	C3: -80.28237E-9
18.56003	0.410751	18.56038	0.00035	
24.30631	0.348969	24.30622	-0.00009	
30.10415	0.293222	30.10389	-0.00026	
34.99336	0.251757	34.99352	0.00015	

Residuals vs. Temperature



Calibration Date: 2021-05-14
Issue Date: 2021-05-17
Calibration ID: 46632

Operator: 
 dлуong

Approver: 
 kmalorny




Certificate of Calibration

Ocean Sonics, Ltd.

Calibration Certificate Number: **C5341**

Test Result:	10 kHz to 100 kHz: -177.9 ± 1.1
	10 kHz to 200 kHz: -178.9 ± 2.1

Model Number	RB9-ETH	Projector Manufacturer	Ocean Sonics
Serial Number	6126	Projector Model	TH2-SER-4F
Manufacture Date	5-Oct-2020	Projector Serial	2225
Measurement Date	8-Oct-2020	Measurement Distance	1 m
Certificate Date	8-Oct-2020	Output Level	130 dB re uPa @ 1 m
Sensitivity @ 250 Hz	-177 dB re V / uPa	Tone Burst	100.0 us / 300 ms
Case Type	R-Type	Reference Manufacturer	Ocean Sonics
Element Manufacturer	Reson	Reference Model	RB9-ETH
Element Model	TC4059-1	Reference Serial	2080
Element Serial	620130	Primary Calibration	20-Jan-2020
Preamp Model	04-300449-01	Preamp Manufacturer	Ocean Sonics
Calibrated By	Cody Ellis	Preamp Model	04-300449-01
Work Order Number	W1234	Preamp Serial	M58
Test Type	RX Sensitivity	Preamp Gain	36 dB
Test Procedure	Complex RMS	ADC Manufacturer	Ocean Sonics
Test Location	Tank #3, 1 m	ADC Model Number	04-300426-01
Water Temperature	16 °C	ADC Serial Number	M59



Frequency kHz	Sensitivity [dBV re uPa]	
	0 deg	90 deg
10.0	-176.7	-176.9
20.1	-177.4	-177.4
30.1	-177.0	-177.1
40.2	-177.2	-177.0
50.2	-177.9	-177.4
60.2	-178.5	-177.8
70.3	-178.4	-178.0
80.3	-178.5	-178.2
90.4	-179.0	-178.4
100.4	-178.6	-178.2
110.4	-178.7	-178.6
120.5	-177.2	-178.9
130.5	-177.1	-179.1
140.5	-177.4	-178.8
150.6	-178.0	-179.0
160.6	-178.2	-179.1
170.7	-178.2	-178.7
180.7	-180.1	-180.3
190.7	-180.3	-180.3
200.0	-181.0	-180.4

Final test checklist



Vangkroken 2
N-1351 RUD
Norway
Tel: +47 6717 4500
Fax: +47 6713 6770
inquiry@nortek-as.com
www.nortek.no

Order number:

45059-1-794

Head ID: AQP 10565
Hardware ID: AQP 16166
Frequency: 1 MHz
Firmware version: 3.43

Harness test
 OK
Type: RS 232 a.in

Recorder erased
 OK
Rec size: 16 GB

Label checked OK
Dock test OK

RS422 cable OK
Length m
Interfacebox 18/48V OK

System type:
Aquadopp Vector Continental
Aquadopp Profiler Awac Other Specify: _____

Tilt check
 Pitch up
 Roll up
 Status bit
 Pitch down
 Roll down
pitch & roll within +/- 0.2°

Heading
 Up
 Down
tolerance: +/- 2°

Pressure
High Accuracy sensor tolerance: +/- 0.1 % of m
Standard sensor tolerance: +/- 0.5 % of m
20 m

 OK

Temperature
 OK
tolerance: +/- 0.1°

Beam	Correct order	Noise floor	Amplitude in tank	Range
Beam 1	<input checked="" type="checkbox"/> OK	10 Counts	> 180 Counts	<input checked="" type="checkbox"/> OK
Beam 2	<input checked="" type="checkbox"/> OK	10 Counts	> 180 Counts	<input checked="" type="checkbox"/> OK
Beam 3	<input checked="" type="checkbox"/> OK	10 Counts	> 180 Counts	<input checked="" type="checkbox"/> OK
Beam 4	<input type="checkbox"/> OK	<input type="text"/> Counts	<input type="text"/> Counts	<input type="checkbox"/> OK

Velocity direction

XYZ coordinate system	ENU coordinate system
X <input checked="" type="checkbox"/> OK	E <input checked="" type="checkbox"/> OK
Y <input checked="" type="checkbox"/> OK	N <input checked="" type="checkbox"/> OK
Z <input checked="" type="checkbox"/> OK	U <input checked="" type="checkbox"/> OK

Head file
 Headfile checked
 Saved as read only

Comments:

External sensors

Power down
 OK

Date
Day 17 Month 11 Year 2020
Fanny Silkevaarden



Pressure Report

Order No:	45059-1-794	Date:	16.11.2020 12:21:54
Head Id:	AQP 10565	Operator:	Faramarz Torkzad
Board Id:	AQD16166	Location:	Nortek factory in Norway
Pressure Range:	20 dBar		

Method: The instrument is put in an automated pressure chamber. Samples are taken at different pressures between 0 dBar and the instrument maximum pressure range. The calibration coefficients are calculated using linear regression with the formula:

$$Pressure = (A0 + A1 * X) / (scaling\ constant)$$

Verification: After the instrument has been updated with the calibration coefficients the instrument is again put in the pressure chamber and sampled at new pressures.

Criteria of acceptance is +/- 0.5% of full scale.

Reference: Paroscientific, Inc. Digiquartz 9000-1K-242. Accuracy 0.01% of 689dBar

Verification

Ref (dBar)	Pressure	Diff	Diff (% of FS)
1.08	1.08	0.01	0.04
2.34	2.36	0.02	0.09
4.38	4.41	0.04	0.18
6.34	6.38	0.04	0.22
7.87	7.92	0.05	0.26
10.74	10.80	0.06	0.30
12.28	12.34	0.06	0.31
14.08	14.14	0.06	0.30
16.23	16.29	0.06	0.28
18.10	18.15	0.05	0.23
19.56	19.59	0.04	0.18

Coefficients

A0	13983
A1	5657
A2	0
A3	0

Nortek AS
Vangkroken 2
NO-1351, Norway
Tel: +4767174500
Fax: +4767136770
inquiry@nortek.no

www.nortek-as.com
True innovation makes a difference

NQM03-400-03 Certificate of Calibration and Test



Tilt Report

Order No: 45059-1-794	Date: 11.11.2020 15:23:26
Head Id: AQP 10565	Operator: Asle Martinsen
Board Id: AQD16166	Location: Nortek factory in Norway

Method: The instrument is fixed in a two axis automated tilt calibration robot that puts the instrument in 5 different positions between -30 degree and + 30 degree for both pitch and roll axes. This is done with the instrument in both up and down orientation. The calibration coefficients are calculated using linear regression with the formula:
 $Tilt\ angle = (A0 + A1 * X + A2 * X^2 + A3 * X^3) / (scaling\ constant)$

Verification: After the instrument has been calibrated with the calibration coefficients, the calibration robot puts the instrument in 5 new positions to verify pitch and roll axes in both up and down orientation.
Criteria of acceptance is +/- 0.2°

Reference: Digital Protractor Series 950 Pro 3600. Accuracy 0.05°

Verification

	Up		Down	
	Pitch	Roll	Pitch	Roll
-30°	-0.03	-0.01	0.00	-0.07
-15°	0.02	0.03	-0.01	-0.03
0°	-0.03	-0.00	0.01	-0.01
15°	-0.01	0.05	0.01	-0.01
30°	0.00	0.01	-0.01	-0.03

Coefficients

	Up		Down	
	Pitch	Roll	Pitch	Roll
A0	18	-284	14	170
A1	13396	13590	13374	13484
A2	-546	1551	-67	-1481
A3	-26180	-27539	-27552	-27981

Nortek AS
 Vangkroken 2
 NO-1351, Norway
 Tel: +4767174500
 Fax: +4767136770
 inquiry@nortek.no

www.nortek-as.com
 True innovation makes a difference

NQM03-400-03 Certificate of Calibration and Test



Compass Report

Order No: 45059-1-794	Date: 11.11.2020 15:35:49
Head Id: AQP 10565	Operator: Asle Martinsen
Board Id: AQD16166	Location: Nortek factory in Norway

Method: The instrument is fixed in a two axis magnetometer calibration jig. The instrument is slowly rotated around the three magnetometer axes. The gain for all axes is calculated and adjusted. Cross correlation between the axes is calculated and adjusted.

Verification: The heading is derived from the magnetometer and tilt sensor measurements. The heading is verified at every 45th degree both with the instrument pointing up and down.
Criteria of acceptance is +/- 2°

Reference: Digital Protractor Series 950 Pro 3600. Accuracy 0.05°

Verification

	Measured (°)		Error (°)	
	Up	Down	Up	Down
0°	1.5	1.3	1.5	1.3
45°	46.5	45.9	1.5	0.9
90°	91.5	90.5	1.5	0.5
135°	136.5	135.8	1.5	0.8
180°	181.4	181.4	1.4	1.4
225°	226.6	226.5	1.6	1.5
270°	271.9	271.7	1.9	1.7
315°	316.9	316.9	1.9	1.9

Transformation matrix

31741	1582	-1466
-135	32767	-101
98	47	32644

Nortek AS
 Vangkroken 2
 NO-1351, Norway
 Tel: +4767174500
 Fax: +4767136770
 inquiry@nortek.no

www.nortek-as.com
 True innovation makes a difference

NQM03-400-03 Certificate of Calibration and Test



Certificate of Calibration and Test

Customer reference: 45059-1-794
Instrument type: Aquapro
Head Id: AQP 10565
Board Id: AQD16166

Calibrations and tests performed

Tilt OK
Compass OK
Pressure OK

Date of approval:

17.11.2020

All the tested values are within Nortek AS specifications

Reviewed and approved by

Nortek AS
Vangkroken 2
NO-1351, Norway
Tel: +4767174500
Fax: +4767136770
inquiry@nortek.no

www.nortek-as.com
True innovation makes a difference

NQM03-400-03 Certificate of Calibration and Test